


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>4110</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>1G Dynamics Limited</h3> <p>Issue No: 008 Issue date: 10 December 2019</p>	
	<p>5 Dunsbridge Business Park Shepreth Royston Herts SG8 6RA</p>	<p>Contact: Graham Hagger Tel: +44 (0)1763 262 112 E-Mail: sales@1g-dynamics.com Website: www.1g-dynamics.com</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
<p>ACCELERATION TRANSDUCERS</p> <p>Reference (precision) Grade</p> <p>Piezoelectric type Transducer at 23°C</p>	<p>High frequency Test Nominal peak acceleration 1g_n up to 10 g_n (9.81 up to 98 m/s⁻²) Charge Sensitivity >0.1 pCg_n (0.01 pC/ms⁻²) <1000 pCg_n (0.01 pC/ms⁻²)</p> <p>20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz</p>	<p>1.5 % 1.3 % 1.8 %</p>	<p>Calibration of charge sensitivity by comparison with a reference (precision grade) transducer</p> <p>Transducer at ambient</p>
<p>Working or non-precision grades Piezoelectric type</p>	<p>High frequency Test Nominal peak acceleration 1g_n up to 10 g_n (9.81 up to 98 m/s⁻²) Charge Sensitivity >0.1 pCg_n (0.01 pC/ms⁻²) <1000 pCg_n (0.01 pC/ms⁻²)</p> <p>20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz</p> <p>Low frequency Test Nominal peak acceleration 0.2 up to 2 g_n (1.96 up to 19.6 m/s⁻²) Charge Sensitivity >2 pCg_n (0.01 pC/ms⁻²) <1000 pCg_n (0.01 pC/ms⁻²)</p> <p>2 Hz to 20 Hz</p>	<p>1.5 % 1.3 % 2.5 % 1.5 %</p>	<p>Calibration of charge sensitivity by comparison with a reference (precision grade) transducer</p> <p>Transducer at ambient</p>



4110
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

1G Dynamics Limited
Issue No: 008 Issue date: 10 December 2019

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
ACCELERATION TRANSDUCERS			Calibration of sensitivity by comparison with a reference (precision grade) transducer
Piezoresistive or strain-gauge type	High frequency Test Nominal peak acceleration 1 up to 10 g _n (9.81 up to 98 m/s ⁻²) System Sensitivity > 1mv/g _n (0.1 mv/ms ⁻²) < 1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
	20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz	1.5 % 1.3 % 1.8 %	
	Low frequency Test Nominal peak acceleration 0.2 up to 2 g _n (1.96 up to 19.6 m/s ⁻²) System Sensitivity >0.05 mv/g _n (0.005 mv/ms ⁻²) < 1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
Integral electronics type	2 Hz to 20 Hz	1.5 %	Calibration of sensitivity by comparison with a reference (precision grade) transducer
	High frequency Test Nominal peak acceleration 1 up to 10 g _n (9.81 up to 98 m/s ⁻²) System Sensitivity > 1 mv/g _n (0.1 mv/ms ⁻²) < 1000 mv/g _n (100 mv/ms ⁻²)		Transducer at ambient
	20 Hz to 5 kHz 5 kHz to 6.3 kHz 6.3 kHz to 10 kHz	1.5 % 1.3 % 1.8 %	



4110
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

1G Dynamics Limited
Issue No: 008 Issue date: 10 December 2019

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
Integral electronics type (Cont'd) PORTABLE ACCELEROMETER CALIBRATORS	Low frequency Test Nominal peak acceleration 0.2 up to 2 g_n (1.96 up to 19.6 m/s^2) System Sensitivity >1 mv/g_n (0.1 mv/ms^{-2}) <1000 mv/g_n (100 mv/ms^{-2}) 2 Hz to 20 Hz Over the ranges detailed above.	1.5 %	Calibration of sensitivity by comparison with a reference (precision grade) transducer Transducer at ambient
END			



4110

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

1G Dynamics Limited

Issue No: 008 Issue date: 10 December 2019

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

Expression of CMCs - symbols and units

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0 μ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %-V + 5.0 μ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10⁻⁶·p·10⁻⁶) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.